


## URGENT: Medical Device Recall Notification

December 21, 2023


Re: Medical Device Safety Recall Notice - ABC Chairs

Please be advised that we have become aware that the Rehabilitation Centre for Children ABC Chair presents a tipping hazard if the footrest is used as a step. We are aware of one occasion where a child using the ABC Chair sustained a minor scrape and the potential for future risk of injury exists. As the Rehabilitation Centre for Children is committed to the utmost safety of our products, we are taking the step of recalling all affected ABC Chairs to eliminate this risk. Please review the figures below to further understand if your product is at risk.




**NO RISK**

Standard Setup with anti-tippers



**NO RISK**

Footrest omitted



**HIGH RISK**

Footrest included without anti-tippers

**If your ABC Chair has a footrest without anti-tippers, we recommend that you discontinue use immediately and contact us.**

If you have an ABC chair in your possession, regardless of footrest set-up, **we are requesting that you contact us upon receipt of this letter to update our records and advise of next steps.**

If the ABC Chair is no longer in use, we are still requesting that you contact us so that we can arrange to pick up the ABC Chair. Any cost incurred in this process will be assumed by the Rehabilitation Centre for Children.

Please contact at (204) 258-6661, or toll-free 1-866-314-0501 or by email at [abcrecall@rccinc.ca](mailto:abcrecall@rccinc.ca)

